



**RADIATION OUTPUT DATA
MX 20**

| <u>KV</u> | <u>R/hr</u> |
|-----------|-------------|
| 10 | 40.0 |
| 11 | 59.2 |
| 12 | 80.4 |
| 13 | 104.8 |
| 14 | 132.8 |
| 15 | 170.4 |
| 16 | 210.6 |
| 17 | 247.0 |
| 18 | 292.6 |
| 19 | 333.0 |
| 20 | 376.1 |
| 21 | 420.2 |
| 22 | 465.2 |
| 23 | 521.6 |
| 24 | 568.6 |
| 25 | 618.2 |
| 26 | 667.4 |
| 27 | 729.4 |
| 28 | 778.8 |
| 29 | 833.0 |
| 30 | 906 |
| 31 | 950 |
| 32 | 1004 |
| 33 | 1062 |
| 34 | 1112 |
| 35 | 1140 |

Technique Factors: Tube Current @ 300 μ A
Source to Probe Distance = 57.2 cm (22.5")
Measuring Instrument: Keithley 35050A dosimeter with Model 96035 probe.
Dose expressed in Roentgens per hour

**225 Larkin Drive, Unit 1
Wheeling, Illinois 60090
Phone: 847-465-9729 Fax: 847-465-9740**